



Short Communication

Detecting endocrine disrupting compounds in water using sulfur-oxidizing bacteria

Steven W. Van Ginkel^{a,1}, Sedky H.A. Hassan^{b,2}, Sang-Eun Oh^{b,*}^a Center for Environmental Biotechnology, Biodesign Institute, Arizona State University, P.O. Box 875701, Tempe, AZ 85287-5701, USA^b Department of Biological Environment, Kangwon National University, Chuncheon, Kangwon-do, South Korea

ARTICLE INFO

Article history:

Received 30 March 2010

Received in revised form 25 May 2010

Accepted 30 May 2010

Available online 1 July 2010

Keywords:

Biosensor

Toxicity

Electrical conductivity

Sulfur-oxidizing bacteria

Endocrine disrupting compounds

ABSTRACT

For the rapid and reliable detection of endocrine disrupting compounds in water, a novel toxicity detection methodology based on sulfur-oxidizing bacteria (SOB) has been developed. The methodology exploits the ability of SOB to oxidize elemental sulfur to sulfuric acid in the presence of oxygen. The reaction results in an increase in electrical conductivity (EC) and a decrease in pH. When endocrine disrupting compounds were added to the system, the effluent EC decreased and the pH increased due to the inhibition of the SOB. We found that the system can detect these chemicals in the 50–200 ppb range, which is lower than many whole-cell biosensors to date. The SOB biosensor can detect toxicity on the order of min to h which can serve as an early warning so as to not pollute the environment and affect public health.

© 2010 Elsevier Ltd. All rights reserved.

1. Introduction

In recent years, more than 100 000 new chemicals of unknown toxicity and effects to humans and wildlife are released into our waterways (Rodriguez-Mozaz et al., 2004; Lean, 2008). The European Commission has stated that 99% of these are not adequately regulated and there is no proper safety information for 95% of them (Lean, 2008). Furthermore, since there are >600 known chemicals listed in the Toxics Release Inventory which could be released, it is difficult to quickly analyse the potential toxicity of each of these chemicals using conventional analytical methods. Hence, the rapid and continuous monitoring of toxicity, in general, is of great importance.

The so-called emerging contaminants include compounds used in everyday life, such as surfactants, pharmaceuticals, plastics, paints, and hormones. These chemicals are of particular concern both because of the volume of these substances used and because of their activity as endocrine disrupting compounds (EDCs) (Rodriguez-Mozaz et al., 2004). When these chemicals enter a wastewater treatment system, their degradation leads to unknown products, such as nonylphenols, which are also produced during the synthesis of polyethoxylate detergents (Farre and Barcelo,

2003). Furthermore, the degradation products of EDCs can be more toxic than the original compounds (Warhurst, 1995).

A wide variety of chemicals can be considered EDCs since they are defined by their biological effect. EDCs have been attributed to cancer, birth defects, declining sperm counts, feminisation of male vertebrates, and other developmental problems in both humans and wildlife (Sadik et al., 2000; Rodriguez-Mozaz et al., 2004; Lean, 2008). With the reuse of water supplies, the concentrations of EDCs in drinking water will likely increase. Even though EDCs in drinking water are not yet regulated in the United States, the US Food and Drug Administration requires environmental risk assessments for new pharmaceutical EDCs with predicted environmental concentrations >1 µg L⁻¹ (Snyder et al., 2000). Bisphenol A, nonylphenol, and diethylstilbestrol are among the chemicals currently being studied as potential EDCs which are partly the focus of this study (Sadik et al., 2000).


According to Hernando et al. (2005), the assessment of contamination of water emerging from contaminated sites or as effluents based on the conventional, analytical detection of specific pollutants is not sufficient to assess environmental risk since toxicity is a biological response (Hernando et al., 2005). Furthermore, the toxicity of pollutants is synergistic meaning the toxicity of multiple pollutants is greater than the sum of toxicity of the individual pollutants (Eaton, 1973; Fernandez-Alba et al., 2002; Hsieh et al., 2006). Therefore, means to evaluate toxicity on living organisms is needed.

Sulfur-oxidizing bacteria (SOB) can be used for monitoring toxicity and their use circumvents many of the obstacles associated with biosensors. SOB, first identified by Sergey Winogradsky in

* Corresponding author. Tel.: +82 33 250 6449; fax: +82 33 241 6640.

E-mail addresses: Steven.vanginkel@asu.edu (S.W. Van Ginkel), sedkyhassan@kangwon.ac.kr (S.H.A. Hassan), ohsangeun@kangwon.ac.kr (S.-E. Oh).¹ Tel.: +1 480 727 9688; fax: +1 480 727 0889.² Tel.: +82 33 243 6449; fax: +82 33 241 6640.

1885, are chemolithotrophic bacteria that oxidize reduced sulfur compounds to sulfuric acid in the presence of oxygen. Some SOB, such as Genus *Thiobacillus*, use elemental sulfur particles as the electron donor, as shown by Eq. (1) (Madigan et al., 2003):



Since elemental sulfur is extremely insoluble, SOB must grow attached to sulfur particles. It is known that extracellular, outer-membrane proteins are considered to play key roles in sulfur mobilization and transport (Rohwerder and Sand, 2003; Zhang et al., 2008). Oxidation of elemental sulfur results in the formation of sulfuric acid (sulfate and hydrogen ion), which lowers the medium pH to ~2.5. The production of sulfate ions also increases the electrical conductivity (EC) of the medium. Generally, EC is proportional to the concentration of ions in the medium and is a measurement of the ability of a medium to carry an electric current and varies both with the number and type of ions the medium contains. Thus, the decrease in pH and the increase in EC reflect the formation of sulfuric acid as the result of sulfur oxidation and can easily be detected using simple EC and pH meters (Madigan et al., 2003). In the presence of toxic chemicals, the activity of SOB will be inhibited, which will cause an increase in pH and a decrease in EC (Oh, 2009). Traditional potentiometric microbial biosensors measure the difference between a working electrode and a reference electrode and the signal is correlated to the concentration of the analyte. This method requires a very stable reference electrode and may be a limitation of these transducers (Farre et al., 2009). However, the SOB biosensor does not have a reference electrode and recalibration is not needed since only changes in pH are needed to detect the inhibition of SOB.

Rogers and Williams note that simplicity, ruggedness (insensitivity to external conditions), and cost-effectiveness are key components of biosensors (Rogers and Williams, 1995). Biosensors should be extremely versatile both in the range of compounds and matrices for which they can be adapted. Biosensors able to detect a variety of chemicals will be most competitive since the market for detecting any single pollutant is small. This work expands

on the work of Oh et al. (2010) which used the SOB-based biosensor to detect a variety of heavy metals in water.

2. Materials and methods

2.1. Culture and medium

Aerobic return activated sludge was used as the inoculum to the SOB biosensors packed with elemental sulfur. The sludge was collected from the Chuncheon Wastewater Treatment plant in Chuncheon City, Kangwondo, Korea. Synthetic stream water was made by diluting the following nutrient mineral buffer (NMB) solution 100 times: NaHCO_3 (3.13 g L^{-1}), NH_4Cl (0.31 g L^{-1}), $\text{NaH}_2\text{PO}_4 \cdot \text{H}_2\text{O}$ (0.75 g L^{-1}), KCl (0.13 g L^{-1}), NaH_2PO_4 (4.22 g L^{-1}), Na_2HPO_4 (2.75 g L^{-1}). Trace metal and vitamin solutions previously described were also diluted 100 times and added to the water (Lovley and Phillips, 1988). The pH, EC, and alkalinity of the synthetic stream water were 6.9 ± 0.2 , 0.12 mS cm^{-1} , and 45 mg L^{-1} as CaCO_3 , respectively, which is similar to the actual stream water which had a pH 6.8–7.0, $\text{EC} = 0.18 \text{ mS cm}^{-1}$, and alkalinity of 75 mg L^{-1} as CaCO_3 .

2.2. Reactor construction and operation

The SOB biosensors tested were two identical cylindrical reactors (100 mL) constructed from acrylic (Fig. 1). Each biosensor was packed with 50 mL of sulfur particles (2–4 mm diameter) and air was introduced from the bottom at a flow rate of 150–250 mL min^{-1} . The biosensors were inoculated with sludge and incubated at 30°C for three d in a batch mode. The biosensors were then fed synthetic stream water continuously in up-flow mode using adjustable peristaltic pumps. The biosensors were operated at a HRT of 30 min for approximately 2–3 d each to reach steady-state conditions (i.e. stable EC and pH values) and then several EDCs were spiked to the influents of the biosensors and the effluent EC and pH values were monitored.




Fig. 1. The SOB biosensor: (a) schematic of the biosensor including sulfur particles, air supply, EC and pH meters. (b) Picture of SOB biosensors used in this study.




Fig. 2. Monitoring of endocrine disruptors by sulfur oxidizing bacterial biosensor: (a) bisphenol A, (b) nonylphenol, (c) estradiol (200 ppb), and (d) diethylstilbesterol (200 ppb).

2.3. Chemicals and analyses

All EDCs were of analytical grade and were used without further purification. The EDCs tested were bisphenol-A, nonylphenol, estradiol, diethylstilbesterol, and tributyltin. All the chemicals used in this study were purchased from Sigma Aldrich. When the EC and the pH of the biosensors reached steady-state conditions (i.e. stable EC and pH values), the EDCs chemicals were then spiked to the influents of the reactors. Toxicity was monitored by measuring changes in the EC (LUTRON meter model # - YK 22 CT) and the pH (LUTRON meter model # - pH 208c). The EC and pH meters were connected to a computer and all data were automatically recorded every 10 min.




Fig. 3. Monitoring of endocrine disruptors by sulfur oxidizing bacterial biosensor: (a) tributyltin, (b) mixture of bisphenol A and nonylphenol (50 ppb each), and (c) mixture of estradiol and diethylstilbesterol (100 ppb each).

3. Results and discussion

The SOB biosensor pH and EC values can be very stable at ~ 2.2 and $\sim 1.1 \text{ mS cm}^{-1}$, respectively, for over 30 d (data not shown). The SOB biosensors were typically operated for 2–3 d before EDCs were added. When the EDCs were added individually at concentrations from 5 to 200 ppb (Figs. 2a–d and 3a), an initial decrease in EC occurred in min to a few h. Tributyltin had the greatest inhibitory effect and decreased the EC to near influent values after 2 h; 100 ppb was much more toxic than 50 ppb.

When bisphenol A and diethylstilbesterol were added individually, the EC reached influent EC values after 50 and 30 h, respectively. The effect due to nonylphenol and estradiol was less pronounced over time reaching about half the in-reactor EC before the chemical was added. When the EDCs were added together (Fig. 3b and c), similar responses occurred. Even though the EDCs were not tested at equal concentrations, tributyltin appears to be the most toxic followed by diethylstilbesterol and bisphenol A, followed by estradiol and nonylphenol. Diethylstilbesterol was more

toxic at 200 ppb (Fig. 2d) than when added at 100 ppb with estradiol also added at 100 ppb (Fig. 3c). Bisphenol A seemed to dominate the toxicity at 50 ppb when added with nonylphenol also at 50 ppb (Figs. 2a–b and 3b).

The initial EC values of the different biosensors shown in Figs. 2 and 3 are slightly different. This is likely due to different sulfur surface areas. Even though 50 mL of sulfur was added to each biosensor, the total mass and surface area of sulfur was different which led to different production rates of sulfuric acid. This phenomenon is not significant since changes in in-reactor EC and pH is what is important for detection. Furthermore, changes in EC are more important than changes in pH since there is low buffering capacity in the influent, and thus even incomplete inhibition will lead to some sulfuric acid and drop the pH into this range. In this regard, it is likely that tributyltin is very toxic and completely inhibited the SOB since the pH increased nearly instantaneously (Fig. 3a).

Although the toxic effects of the EDCs tested in this study are still somewhat unclear, we can make some conclusions on their effects on SOB. Some known EDCs cause cellular toxicity related to DNA damage, protein damage, oxidative damage, and membrane damage due to fatty acid limitation. In humans, these effects are in addition to estrogenic effects (Omowunmi and Diane, 1999; Gu et al., 2002). Gu et al. (2002) suggests bisphenol A may cause DNA damage during replication while nonylphenol may interfere with respiration after partitioning into the cellular membrane. Similarly, tributyltin which has high lipophilicity and also a protonophore may also partition into cellular membranes and interfere with respiration (Bragadin et al., 1999; Martins et al., 2005). At the low biosensor pH of 2–3, this kind of inhibitory effect on respiration, i.e. the proton motive force, is enhanced. DNA damage or the inability to replicate DNA may not be the main mode of inhibition since these EDCs are detected on the order of hs – a period of time shorter or equal to what the cells need to replicate.

There still is a lack of systems suitable for determination of these EDCs. Sadik et al. (2000) using cyclic voltammetry observed limits of detection of bisphenol A, nonylphenol, estradiol and diethylstilbestrol in the 0.5–1 mM range (~100–300 ppm) which is a detection limit an order of magnitude higher than reported here (Sadik et al., 2000). Using *V. fischeri*, Farre and Barcelo observed an EC₅₀ of 360 ppb for nonylphenol (Farre and Barcelo, 2003). Using a bioluminescent *E. coli* strain, Min et al. (1999) observed a minimum detection level for bisphenol A at 500 ppb – an order of magnitude higher than reported here. Warhurst states that alkylphenol ethoxylates have LC₅₀s above about 1.5 ppm, whereas nonylphenol LC₅₀s are generally around 100 ppb (Warhurst, 1995). Here, we detected nonylphenol at 50 ppb in a matter of h. Since some rivers in Switzerland contain alkylphenolic metabolite concentrations on the order of tens of ppb, our detection limit should be lower (Warhurst, 1995).

4. Conclusions

A total of five EDCs were detected on the order of min to h in the 50–200 ppb range which is lower than many whole-cell biosensors to date. Furthermore, there are several advantages of using the SOB-based biosensor which circumvent many of the obstacles associated with biosensors. First, it is simple, uses little energy, and can be operated for at least several months. The air sparging of sulfur particles laden with SOB in a continuous flow reactor while measuring the pH and EC is sufficient for detecting toxicity. Second, periodic calibration of pH and EC meters are not needed

because EC or pH changes over time are more important than exact values. Third, sulfur particles are very cheap and no toxic byproducts are produced – the total mass of sulfate produced is low. Fourth, the SOB-based biosensor is versatile in that it is able to detect a variety of toxic chemicals at the same time.

Acknowledgements

This work was supported by the 2007 Eco-star project funded by the Korean Ministry of Environment and by the Institute of Environmental Research at KNU.

References

- Bragadin, M., Marton, D., Iero, A., Manente, S., Perin, G., Rizzoli, V., Scutari, G., 1999. A bioanalytical method for the monitoring of metal alkyls in solution. *Anal. Biochem.* 269, 420–423.
- Eaton, J.G., 1973. Chronic toxicity of a copper, cadmium and zinc mixture to the fathead minnow (*Pimephales promelas rafinesque*). *Water Res.* 7, 1723–1736.
- Farre, M., Barcelo, D., 2003. Toxicity testing of wastewater and sewage sludge by biosensors, bioassays and chemical analysis. *Trends Anal. Chem.* 22, 299–310.
- Farre, M., Kantiani, L., Perez, S., Barcelo, D., 2009. Sensors and biosensors in support of EU directives. *Trends Anal. Chem.* 28, 170–185.
- Fernandez-Alba, A.R., Hernandez Guil, M.D., Lopez, G.D., Chisti, Y., 2002. Comparative evaluation of the effects of pesticides in acute toxicity luminescence bioassays. *Anal. Chim. Acta* 451, 195–202.
- Gu, M.B., Min, J., Kim, E.J., 2002. Toxicity monitoring and classification of endocrine disrupting chemicals (EDCs) using recombinant bioluminescent bacteria. *Chemosphere* 46, 289–294.
- Hernando, M.D., Fernandez-Alba, A.R., Tauler, R., Barcelo, D., 2005. Toxicity assays applied to wastewater treatment. *Talanta* 65, 358–366.
- Hsieh, S.H., Tsai, K.P., Chen, C.Y., 2006. The combined toxic effects of nonpolar narcotic chemicals to *Pseudokirchneriella subcapitata*. *Water Res.* 40, 1957–1964.
- Lean, G., 2008. It's Official: Men Really are the Weaker Sex. *The Independent*. <<http://www.independent.co.uk/news/science/its-official-men-really-are-the-weaker-sex-1055688.html>>.
- Lovley, D.R., Phillips, E.J.P., 1988. Novel mode of microbial energy metabolism: organic carbon oxidation coupled to dissimilatory reduction of iron or manganese. *Appl. Environ. Microbiol.* 54, 1472–1480.
- Madigan, M.T., Martinko, J.M., Parker, J., 2003. *Brock Biology of Microorganisms*. Pearson Education Inc., New York.
- Martins, J.D., Jurado, A.S., Moreno, A.J.M., Madeira, V.M.C., 2005. Comparative study of tributyltin toxicity on two bacteria of the genus *Bacillus*. *Toxicol. in Vitro* 19, 943–949.
- Min, J., Kim, E.J., LaRossa, R.A., Gu, M.B., 1999. Distinct responses of a *recA::luxCDABE* *Escherichia coli* strain to direct and indirect DNA damaging agents. *Mut. Res.* 442, 61–68.
- Oh, S.E., 2009. *Water Toxicity Detecting Apparatus and Method Using Sulfur Particles*. University–Industry Cooperation Foundation, Kangwon University, USA.
- Oh, S.E., Hassan, S.H.A., Van Ginkel, S.W., 2010. A novel biosensor for detecting toxicity in water using sulfur-oxidizing bacteria. *Sens. Actuators B Chem.* doi:10.1016/j.snb.2010.01.052.
- Omowunmi, A.S., Diane, M.W., 1999. Novel monitoring strategies for endocrine disrupting chemicals. *Environ. Sci. Technol.* 33, 368A–374A.
- Rodriguez-Mozaz, S., Marco, M.P., Lopez De Alda, M.J., Barcelo, D., 2004. Biosensors for environmental monitoring of endocrine disruptors: a review article. *Anal. Bioanal. Chem.* 378, 588–598.
- Rogers, K.R., Williams, L.R., 1995. Biosensors for environmental monitoring: a regulatory perspective. *Trends Anal. Chem.* 14, 289–294.
- Rohwerder, T., Sand, W., 2003. The sulfane sulfur of persulfides is the actual substrate of the sulfur-oxidizing enzymes from *Acidithiobacillus* and *Acidiphilium* spp. *Microbiology* 149, 1699–1709.
- Sadik, O.A., Ngundi, M.M., Yan, F., 2000. Environmental biosensors for organochlorines, cyanobacterial toxins and endocrine disrupting chemicals. *Biotechnol. Bioprocess Eng.* 5, 407–412.
- Snyder, E.M., Pleus, R.C., Snyder, S.A., 2000. *American Water Works Association*, vol. 97. <<http://www.awwa.org/files/Membership/SectionDownloads/Pharm.doc>>.
- Warhurst, A.W., 1995. *An Environmental Assessment of Alkylphenol Ethoxylates and Alkylphenols*. Friends of the Earth, London.
- Zhang, C.G., Zhang, R.Y., Xia, J.L., Zhang, Q., Nie, Z.Y., 2008. Sulfur activation-related extracellular proteins of *Acidithiobacillus ferrooxidans*. *Trans. Nonferr. Metals Soc. (English Edition)* 18, 1398–1402.