

Folding-Based Electrochemical Biosensors: The Case for Responsive Nucleic Acid Architectures

ARICA A. LUBIN[†] AND KEVIN W. PLAXCO^{*,†,‡}

[†]Department of Chemistry and Biochemistry, [‡]Biomolecular Science and Engineering Program, University of California, Santa Barbara, California 93106

RECEIVED ON JUNE 4, 2009


CON SPECTUS

Biomolecular recognition is versatile, specific, and high affinity, qualities that have motivated decades of research aimed at adapting biomolecules into a general platform for molecular sensing. Despite significant effort, however, so-called “biosensors” have almost entirely failed to achieve their potential as reagentless, real-time analytical devices; the only quantitative, reagentless biosensor to achieve commercial success so far is the home glucose monitor, employed by millions of diabetics.

The fundamental stumbling block that has precluded more widespread success of biosensors is the failure of most biomolecules to produce an easily measured signal upon target binding. Antibodies, for example, do not change their shape or dynamics when they bind their recognition partners, nor do they emit light or electrons upon binding. It has thus proven difficult to transduce biomolecular binding events into a measurable output signal, particularly one that is not readily spoofed by the binding of any of the many potentially interfering species in typical biological samples. Analytical approaches based on biomolecular recognition are therefore mostly cumbersome, multistep processes relying on analyte separation and isolation (such as Western blots, ELISA, and other immunochemical methods); these techniques have proven enormously useful, but are limited almost exclusively to laboratory settings.

In this Account, we describe how we have refined a potentially general solution to the problem of signal detection in biosensors, one that is based on the binding-induced “folding” of electrode-bound DNA probes. That is, we have developed a broad new class of biosensors that employ electrochemistry to monitor binding-induced changes in the rigidity of a redox-tagged probe DNA that has been site-specifically attached to an interrogating electrode. These folding-based sensors, which have been generalized to a wide range of specific protein, nucleic acid, and small-molecule targets, are rapid (responding in seconds to minutes), sensitive (detecting sub-picomolar to micromolar concentrations), and reagentless. They are also greater than 99% reusable, are supported on micrometer-scale electrodes, and are readily fabricated into densely packed sensor arrays. Finally, and critically, their signaling is linked to a binding-specific change in the physics of the probe DNA, and not simply to adsorption of the target onto the sensor head. Accordingly, they are selective enough to be employed directly in blood, crude soil extracts, cell lysates, and other grossly contaminated clinical and environmental samples. Indeed, we have recently demonstrated the ability to quantitatively monitor a specific small molecule in real-time directly in microliters of flowing, unmodified blood serum.

Because of their sensitivity, substantial background suppression, and operational convenience, these folding-based biosensors appear potentially well suited for electronic, on-chip applications in pathogen detection, proteomics, metabolomics, and drug discovery.



Introduction

Biomolecular recognition is second to none when it comes to affinity, specificity, and breadth. As a result, analytical approaches based on this phenomenon, including Western blots, ELISAs, and other immunochemical methods, dominate molecular pathology.^{1–3} These approaches, however, remain cumbersome, multistep, laboratory-bound *processes*. Indeed, the only commercially viable biosensing *device* (as opposed to a *process*) on the market today is the glucose-oxidase-based blood sugar sensor, first described in the 1960s^{4,5} before coming into widespread use by diabetics across the globe.

A single, fundamental problem has largely thwarted the development of effective, reagentless biosensors: most biomolecules do not produce any readily measurable signal upon target binding. An antibody, for example, neither changes its shape or dynamics nor emits light or electrons upon binding its antigen. How, then, does one build a sensor based on biological recognition? A common approach in the field has, historically, been to tie the recognition biomolecule to a surface and then measure the change in refractive index (e.g., surface plasmon resonance^{6,7}), mass (e.g., quartz crystal microbalance⁸), steric bulk (e.g., static microcantilever,^{9,10} impedance spectroscopy^{11,12}), or charge (e.g., field-effect-transistor-based sensors^{13,14}) that occurs when the target (typically a bulky, charged macromolecule) is bound. Unfortunately, while several of these adsorption-based methods, such as the surface-plasmon resonance approach commercialized by Biacore,^{15–17} have proven of utility in academic laboratories, where high purity samples are the norm, they have also proven easily foiled by false positives in complex contaminant-ridden “real” samples. Studies of human blood serum, for example, have identified nearly 3000 distinct proteins to date,¹⁸ and it is estimated that the 60–80 mg total protein content of this material consists of up to 10 000 distinct components.^{19–21} Since these proteins also have mass, charge, and so on, the nonspecific adsorption of any of them is difficult or impossible to distinguish from the binding of the authentic target [e.g., ref 12].

Given that the field of marketable biosensors is so nascent, it is informative to explore the principles underlying the success of the single commercially viable, quantitative biosensor on the market to date: the home glucose meter. Key to its success is the fact that, unlike most biomolecules, this sensor’s recognition element, glucose oxidase, produces a *readily measurable response* when it binds its target. Specifically, when it binds (and chemically transforms) its target, the enzyme produces hydrogen peroxide, which is then detected electrochem-

ically. That is, nature has given us a gift with this protein: it has produced a mechanism that transduces target binding into a specific, readily detected output not easily spoofed by the non-specific adsorption of other materials to the sensor surface. This observation lies at the heart of the approach we have taken to the design of biosensors. Namely, the key development will be the identification of mechanisms linking biomolecular recognition with large-scale physical changes which, in turn, can be transduced into specific output signals.

Some biomolecules fold only upon binding their complementary target, thus linking recognition with an enormous change in conformation and dynamics [e.g., refs 22 and 23]. For example, single-stranded DNA is unfolded (flexible, dynamic, and fully solvated) in the absence of its “target” (its complementary strand) but “folds” into a rigid, well-structured double helix upon hybridization. More generally, it appears that most any single-domain protein or nucleic acid can be re-engineered such that it adopts its native conformations only upon recognizing its binding partner.^{24,25} That is, because biomolecular folding is generally cooperative, it is easy to identify mutations that push the folding equilibrium heavily toward unfolding. Even largely unfolded mutants, however, still sample the native state, and, as only the native state binds the target, the presence of the target rapidly drives the equilibrium back to this state. That these equilibrium-unfolded biomolecules couple recognition with the least subtle of all possible conformational changes (folding itself) suggests that this effect can serve as a robust and general signal transduction mechanism.

From Signal Transduction to Biosensor: E-DNA

Signal transduction alone, however, does not make a biosensor: the binding event must also be transduced to an output signal that can be detected unambiguously against the background arising from any other molecules (interferants) likely in the sample. To this end, a number of optical reporters of binding-induced folding have been reported, [e.g., refs 24–33]. For example, Kramer, a pioneer in this field, has developed fluorescent reporters that have been widely used to detect the binding (hybridization) induced “unfolding” of stem-loop DNA structures, commonly known as “molecular beacons”.³¹ Likewise, we and others have demonstrated optical reporters for the binding-induced folding of polypeptides, proteins, and nucleic acid aptamers.^{25–27} From PNA-peptide chimera beacons to protein molecular switches to RNA aptamer beacons, these optical sensors detect targets ranging from antibodies to small molecules in the laboratory.




FIGURE 1. The first E-DNA sensor comprised a redox-tagged stem-loop DNA attached to an interrogating electrode.³⁴ In the absence of target, the redox tag is held in proximity to the electrode, ensuring efficient electron transfer (eT) and a large, readily detectable faradaic current. Upon hybridization with a target, the redox tag is removed from the electrode, impeding the signaling current.

Unfortunately, however, fluorescent and strongly absorbing interferants are common in clinical and environmental samples, which, in practice, reduces the utility of such optical approaches in “real-world” settings [e.g., ref 28]. Electroactive interferants, in contrast, are relatively rare, suggesting that electrochemical approaches to monitoring binding-induced folding might be of broader utility. This idea, binding-induced biomolecular folding monitored electrochemically, has been the dominant focus of our research program.

The first electrochemical biosensor based on binding-induced folding was developed in 2003 by Chunhai Fan, then a postdoc in our group. As proof of principle, he immobilized one end of a stem-loop DNA onto a gold electrode via a terminal thiol modification (forming a self-assembled monolayer) with the other end bearing a redox-reporting ferrocene^{34,35} to create an electrochemical DNA (E-DNA) sensor analogous to the optical stem-loop molecular beacons of Kramer.³¹ In the absence of target, the stem structure of the probe DNA holds the redox tag in proximity to the electrode, producing a large faradaic current. Hybridization to a complementary oligonucleotide removes the redox tag from proximity to the electrode, significantly decreasing this current and supporting the ready detection of specific DNA sequences (Figure 1). Given the potential advantages of electrochemical detection (which we detail below), and the close analogy between E-DNA sensors and molecular beacons, it is perhaps not surprising that, within less than a year, several other groups reported them independently reported similar sensors,^{36,37} and in the years since numerous other authors have described their fabrication, characterization, optimization, and applications.³⁸

The E-DNA platform exhibits a number of potentially appealing attributes.^{39–42} First, it achieves sub-nanomolar detection limits, discriminates between at least modestly mismatched targets, and responds rapidly (equilibration half-lives of between <1 and 30 min for targets of 17 to 100 bases).^{34,43–45} Second, because E-DNA signaling is associ-

ated with a ligand-induced change in the conformation and dynamics of the DNA probe and not simply absorption of the target to the sensor, the approach is quite impervious to complex “real” sample conditions and performs well even when challenged directly in multicomponent, contaminant-ridden samples such as blood serum, saliva, urine, foodstuffs, and soil suspensions (Figure 2).^{44–46} Third, because the E-DNA probe is strongly chemi-adsorbed onto its electrode, the approach is reagentless and readily reusable (we recover >99% of our signal after a simple, aqueous wash⁴⁴). Finally, because E-DNA is electrochemical, it is both scalable (we regularly employ micrometer-scale electrodes) and amenable to parallelization and device integration.^{47,48}

Sensor Optimization and the Mechanism of E-DNA Sensing

In the years since the inception of E-DNA, we have characterized a number of fabrication and operational parameters in an attempt to improve the signaling, reproducibility, and shelf life of the platform. Specifically, we have monitored the effects of varying probe length,⁴⁹ the nature of the coadsorbate forming the self-assembled monolayer,^{46,50–52} the density with which the probe DNAs are packed on the sensor and the electrochemical method employed⁵³ on E-DNA performance and have achieved significant improvements in sensor gain and shelf life. These efforts have also shed light on the platform’s signaling mechanism. Specifically, studies of the effects of probe packing density demonstrate that when the mean spacing between probes is less than their length ($\sim 10^{12}$ probe molecules/cm²) the signal change upon hybridization *improves* with increasing probe density, presumably as crowding effects between the neighboring probe-target duplexes minimize electron transfer from the bound state and thus increase the observed signal change (Figure 3). This suggests, in turn, that E-DNA signaling arises due to binding-linked changes in the efficiency with which the terminal redox tag strikes the elec-




FIGURE 2. E-DNA sensors are rapid, selective, and readily reusable.⁴⁵ (Top) For example, an E-DNA sensor specifically responds to target even in the presence of 50 000× excess genomic DNA and when deployed directly in complex “dirty” samples such as clinical materials, soil suspensions, and foodstuffs. (Bottom) E-DNA sensors equilibrate within minutes and are regenerated via a simple, 30 s wash with deionized water.




FIGURE 3. The signal gain of E-DNA sensors is a function of the density with which the probe DNAs are packed onto the electrode surface.⁵³ Specifically, signaling *improves* with increasing probe density, presumably because crowding between neighboring probe-target duplexes minimizes electron transfer from the bound state, resulting in increased signal change upon hybridization.

trode (i.e., with collision dynamics) and not to the binding-induced conformational change per se.⁵³

Two additional lines of evidence support the “collision efficiency model” of E-DNA signaling. First, both our group and that of Inouye have observed that gain of E-DNA sensors can

be tuned by varying the frequency at which the potential is modulated in alternating current or square wave voltammetry.^{53–55} Second, E-DNA sensors fabricated using single-stranded, linear probes *also* support efficient E-DNA signaling, presumably because the flexible, single-stranded state




FIGURE 4. This “signal-on” E-DNA architecture, based on a target-induced strand displacement mechanism, achieves sub-picomolar detection limits.⁵⁷ In this mechanism, sensor target binding displaces a flexible, single-stranded element modified with a redox tag. This, in turn, strikes the electrode, generating a large increase in faradaic current.

also supports efficient collisions.^{53,55} Indeed, because target binding no longer competes with the unfavorable energy of breaking the stem, the gain of linear probe sensors is improved relative to the equivalent stem-loop architecture.⁴⁹

Increasing Signal Gain: From Signal-off to Signal-on Architectures

A limitation of the first E-DNA architecture is that it is a “signal-off” sensor in which target recognition is signaled by the loss of an initially high current. A drawback is that this limits signal gain: one can never suppress more than 100% of the original current. Conversely, “signal-on” sensors, in theory, can approach enormous signal gain as the background observed in the absence of target is pushed toward zero. (By analogy fluorescence spectroscopy, a low-background “signal-on” technology, is much more sensitive than absorbance spectroscopy, which is “signal-off”). Thus motivated we, and others, have explored several signal-on E-DNA architectures.^{56–59}

The first signal-on E-DNA sensor design, described by Immoos et al., consisted of a DNA–polyethylene glycol–DNA triblock probe immobilized onto an electrode and labeled at its distal terminus with a redox reporter.⁵⁶ Hybridization with a target that is complementary to both of the flanking DNA elements brings the labeled end of the probe into proximity with the electrode, increasing electron transfer. This design results in 600% gain and a detection limit of ~ 200 pM.⁵⁶ In an effort to improve on this, we developed a second signal-on sensor via a strand displacement mechanism in which target binding displaces a flexible, single-stranded element modified with a redox tag. Upon target binding, the redox molecule at the terminus of the displaced strand generates up to a 700% increase in faradaic current, supporting detection down to sub-picomolar target concentrations⁵⁷ (Figure 4). A complication of the strand-displacement strategy, however, is

that the labeled probe, which is the probe that is partially displaced upon target binding, is only held onto the sensor via hybridization with the surface attached strand and thus the stability and reusability of this sensor is poor. In response, we have more recently described a construct comprising a single DNA chain that, in the absence of target, adopts a double-stem-loop pseudoknot structure that holds the redox tag away from the surface.^{58,59} Target binding disrupts the pseudoknot, liberating a flexible, single-stranded element that more readily collides with the electrode surface and produces a strong signal increase.

The Detection of Non-Nucleic Acid Targets

The above studies suggest that the only requisite for E-DNA signaling is that binding alters the flexibility of the signaling probe. This, in turn, suggests that the approach can be expanded to the detection of other targets provided they specifically bind to our probe and, in turn, alter its collision dynamics. To this end, Yi Xiao in our laboratory introduced aptamers, DNA or RNA sequences selected *in vitro* for specific and high affinity binding to a given target molecule,^{60–62} into the E-DNA platform.⁶³ Critically, aptamers can be re-engineered such that they only fold upon target binding, thus coupling recognition with the type of large-scale change in flexibility required to generate an E-DNA-like signal [e.g., refs 64–67].

The first of these electrochemical, aptamer-based (E-AB) sensors utilized a 31-base probe that binds the enzyme thrombin.^{64,68–70} Following on the E-DNA platform, we labeled the 3' terminus of this probe with a redox label and attached its 5' terminus to a gold electrode (Figure 5). This system produces a readily measurable decrease in faradaic current upon binding as little as 20 nM thrombin and is selective enough to deploy directly in blood serum.⁷¹ This first E-AB




FIGURE 5. The first electrochemical, aptamer-based (E-AB) sensor comprised a redox-tagged DNA aptamer directed against the blood-clotting enzyme thrombin.⁷¹ Thrombin binding reduces the current from the redox tag, readily signaling the presence of the target.




FIGURE 6. A signal-on E-AB sensor based on the strand-displacement mechanism and directed against the protein thrombin achieves a 10-fold increase in signal gain over its signal-off counterpart (Figure 5) and, in turn, a 7-fold increase in sensitivity.⁷²

sensor, however, was signal-off and low gain. In an effort to convert it into a higher-gain, signal-on sensor, Xiao later employed a strand-displacement mechanism analogous to the equivalent signal-on E-DNA sensor (Figure 6). This new sensor yielded a $\sim 300\%$ increase in signal at saturating target and a detection limit of just 3 nM.⁷²

As was the case with our earlier E-DNA sensor, other reports quickly followed. For example, just a few months after our publication, O'Sullivan and co-workers described a sensor based on a similar aptamer sequence in which they observed an increase in faradaic current upon thrombin binding and a detection limit of 500 pM.^{73,74} Since then, other E-AB sensors have been reported against a wide range of targets [e.g., refs 75–84]. Employing a cocaine-binding aptamer, for example, we developed an E-AB sensor that can detect low micromolar concentrations of this target directly in blood serum and other complex sample matrices.^{75,76} Indeed, because the on- and off-rates for cocaine binding are rapid, this sensor supports real-time detection directly in flowing, undiluted blood serum (Figure 7).⁷⁷ O'Sullivan and co-workers then exploited the observation that the thrombin-binding aptamer also folds in

the presence of potassium to detect this important ion,⁷⁸ and we reported an E-AB sensor directed against the protein platelet-derived growth factor that achieves a 50 pM detection limit directly in blood serum.⁷⁹ More recently still, Qu and co-workers have developed an E-AB sensor directed against carbon nanotubes⁸⁰ and Kim and co-workers described a mercury sensor with a detection limit of 100 nM.⁸¹ Expanding the approach by employing RNA, rather than DNA, aptamers, Gothelf and co-workers have developed a sensor for the asthma drug theophylline⁸² that achieves nanomolar detection limits in buffer⁸² and a somewhat lower sensitivity in diluted, RNase-inhibited serum.^{83,84}

All of the above sensors utilize probes that are contiguous strands attached to the interrogating electrode. However, in an approach similar to our strand-displacement E-AB sensor, several groups have used binding-induced folding to completely displace an aptamer from an electrode-bound complementary strand⁸⁵ (or a complementary strand from an electrode-bound aptamer⁸⁶) to generate a faradaic signal and detect ATP. Likewise, we have utilized the strand cleaving ability of




FIGURE 7. Small molecule E-AB sensors support real-time detection even in complex sample matrices, such as blood serum. The cocaine E-AB sensor, for example, supports real-time detection of cocaine directly in undiluted blood serum as it flows through a submicroliter chamber.⁷⁷ The letter designations on the external leads denote the “reference” (R), “counter” (C), and “working” (W) electrodes to which they are attached.




FIGURE 8. Utilizing double-stranded DNA as a support scaffold for a small molecule receptor, we have fabricated E-DNA-like sensors for the detection of protein–small-molecule interactions.⁸⁹ Shown, for example, is the detection of low nanomolar concentrations of antibodies against the drug digoxigenin.

a lead-dependent DNAzyme to detect parts per billion lead directly in soil extracts.⁸⁷

The Detection of DNA Binding Proteins

E-AB sensors expanded our approach to the detection of specific proteins and small molecules, but aptamers are not the only oligonucleotides that bind nonoligonucleotide targets. For example, while on sabbatical in our laboratory, Francesco Ricci of the University of Rome, Tor Vergata employed double-stranded or single-stranded DNA probes to detect several specific double- and single-stranded binding proteins at concentrations as low as 2 nM.⁸⁸ Sensors employing single-stranded DNA probes, for example, supported the detection of endogenous single-strand binding activity in minutes, directly in crude cellular extracts.⁸⁸

Non-Nucleotide Recognition Elements

While the E-DNA/E-AB platform is versatile, for several years it remained limited to analytes that bind nucleic acids. Recognizing this, Ricci and Cash in our group set out to expand the approach by appending a small molecule recognition element onto a double-stranded DNA probe that acts as a phys-

ical scaffold (Figure 8).⁸⁹ In the absence of target, a redox tag appended to this scaffold collides with the electrode, permitting electron transfer. The binding of a macromolecule to the receptor alters the efficiency of these collisions, leading to a change in faradaic current.⁸⁹ First developed with the receptors biotin and digoxigenin, this technology readily detects sub-nanomolar to low nanomolar concentrations of antibodies and other small-molecule-binding proteins in complex, contaminant-ridden samples such as blood and blood serum. Recently, for example, it has been employed to detect both anti-TNT antibodies and, via a competition assay, parts-per-trillion TNT in crude soil extracts.⁹⁰ Moving forward, we believe this “E-DNA scaffold” approach may prove a general method for the detection of macromolecules that bind to low molecular weight targets for application in, for example, drug screening.

Conclusion and Outlook

The E-DNA platform is sensitive, convenient, specific, and, importantly, selective enough to deploy directly in whole blood, cellular lysates, soil extracts, and other realistically complex samples. Moreover, the approach requires only that tar-

get binding alter the efficiency with which the redox tag on the probe biomolecule strikes the electrode, which can be induced by a switch between two discrete conformational states (as in the original stem-loop E-DNA sensor, or the various "strand-displacement" architectures we have described), by wholesale folding (as in E-AB sensors and linear-probe E-DNA sensors) or via a change in steric bulk, charge, or hydrodynamic radius (for the detection of DNA-binding proteins and in the E-DNA scaffold approach). This, in turn, lends the approach great versatility and accounts for the wide range of targets for which such sensors have already been reported.

Given the potential promise of folding-based electrochemical biosensors, what does the future hold? Unfortunately, the majority of the sensors described here are directed against analytes notable more for their convenience as test-bed targets than their value as clinically relevant diagnostics. This has largely precluded the rigorous side-by-side comparison of these approaches with current gold-standard clinical methods on authentic clinical samples. Thus, while the promise of this platform appears strong, it remains to be seen if it can deliver the clinically relevant specificity and sensitivity required to be of widespread use. Nevertheless, given the approach's many positive attributes, and given the growing number of research groups working on improving and expanding it, we are increasingly optimistic that platforms similar to those described in this Account will offer viable solutions to a wide range of analytical problems.

We gratefully acknowledge our colleagues and collaborators, with particular recognition going to Chunhai Fan and our long-standing collaborators Alan Heeger and Yi Xiao. We also thank Kevin Cash, Brian Baker, Rebecca Lai, Norbert Reich, Francesco Ricci, H. Tom Soh, James Sumner, James Swensen, and Ryan White. This work was supported by the NIH (GM062958-01 and 2R01EB002046) and the Institute for Collaborative Biotechnologies (DAAD19-03-D-0004 from the U.S. Army Research Office).

BIOGRAPHICAL INFORMATION

Arica Lubin received her B.S. in Biochemistry from UCSB in 2000. Continuing her education, she began her Ph.D. studies in Biochemistry under Plaxco in 2002 and finished her degree in 2009. Her work has focused on the development and characterization of electrochemical DNA sensors.

Kevin Plaxco completed his Ph.D. studies performing molecular dynamics simulations under Goddard at Caltech. Following stints in experimental molecular biophysics with Dobson at Oxford and Baker at Washington, he came to UCSB in 1998 and set up a group focused on protein folding. Realizing the poten-

tial engineering applications of folding, he shifted gears once again in the mid-2000s to focus on folding-based detection and responsive materials and devices.

FOOTNOTES

*To whom correspondence should be addressed: E-mail: kwp@chem.ucsb.edu. Phone: (805) 893-5558 Fax: (805) 893-4120.

REFERENCES

- Engvall, E.; Perlman, P. Enzyme-linked immunosorbent assay (ELISA). Quantitative assay of immunoglobulin G. *Immunochemistry* **1971**, *8*, 871–874.
- Van Weemen, B. K.; Schuur, A. H. Immunoassay using antigen-enzyme conjugates. *FEBS Lett.* **1971**, *15*, 232–236.
- Renart, J.; Reiser, J.; Stark, G. R. Transfer of proteins from gels to diazobenzyloxymethyl-paper and detection with antisera: a method for studying antibody specificity and antigen structure. *Proc. Natl. Acad. Sci. U.S.A.* **1979**, *76*, 3116–3120.
- Clark, L. C.; Lyons, C. Electrode Systems for Continuous Monitoring in Cardiovascular Surgery. *Ann. N.Y. Acad. Sci.* **1962**, *102*, 29–45.
- Updike, S. J.; Hicks, G. P. The enzyme electrode. *Nature* **1967**, *214*, 986–988.
- Gouzy, M.; Keß, M.; Krämer, P. M. A SPR-based immunosensor for the detection of isoproterenol. *Biosens. Bioelectron.* **2008**, *24*, 1563–1568.
- Rich, R. L.; et al. A global benchmark study using affinity-based biosensors. *Anal. Biochem.* **2009**, *386*, 194–216.
- Frisk, T.; Sandstrom, N.; Eng, L.; van der Wijngaart, W.; Månsson, P.; Stemme, G. An integrated QCM-based narcotics sensing microsystem. *Lab Chip* **2008**, *8*, 1648–1657.
- Shu, W.; Laurenson, S.; Knowles, T. P. J.; Ferrigno, P. K.; Seshia, A. A. Highly specific label-free protein detection from lysed cells using internally referenced microcantilever sensors. *Biosens. Bioelectron.* **2008**, *24*, 233–237.
- Hansen, K. M.; Thundat, T. Microcantilever biosensors. *Methods* **2005**, *37*, 57–64.
- Du, Y.; Li, B.; Wei, H.; Wang, Y.; Wang, E. Multifunctional label-free electrochemical biosensor based on an integrated aptamer. *Anal. Chem.* **2008**, *80*, 5110–5117.
- Bogomolova, A.; Komarova, E.; Reber, K.; Gerasimov, T.; Yavuz, O.; Bhatt, S.; Aldissi, M. Challenges of Electrochemical Impedance Spectroscopy in Protein Biosensing. *Anal. Chem.* **2009**, *81*, 3944–3949.
- Sakata, T.; Miyahara, Y. Potentiometric detection of single nucleotide polymorphism by using a genetic field-effect transistor. *ChemBioChem* **2005**, *6*, 703–710.
- Park, H. J.; Kim, S. K.; Park, K.; Lyu, H. K.; Lee, C. S.; Chung, S. J.; Yun, W. S.; Kim, M.; Chung, B. H. An ISFET biosensor for the monitoring of maltose-induced conformational changes in MBP. *FEBS Lett.* **2009**, *583*, 157–162.
- Fägerstam, L. G.; Frostell, Å.; Karlsson, R.; Kullman, M.; Larsson, A.; Malmqvist, M.; Butt, H. Detection of antigen-antibody interactions by surface plasmon resonance. Application to epitope mapping. *J. Mol. Recognit.* **1990**, *3*, 208–214.
- Fivash, M.; Towler, E. M.; Fisher, R. J. BIAcore for macromolecular interaction. *Curr. Opin. Biotechnol.* **1998**, *9*, 97–101.
- Rich, R. L.; Myszyka, D. G. Survey of the year 2007 in commercial optical biosensor literature. *J. Mol. Recognit.* **2008**, *21*, 355–400.
- Tang, H.; Ali-Khan, N.; Echan, L. A.; Levenkova, N.; Rux, J. J.; Speicher, D. W. A novel four-dimensional strategy combining protein and peptide separation methods enables detection of low-abundance proteins in human plasma and serum proteomes. *Proteomics* **2005**, *5*, 3329–3342.
- Wrotnowski, C. The future of plasma proteins. *Genet. Eng. News* **1998**, *18*, 14.
- Adkins, J. N.; Varnum, S. M.; Auberry, K. J.; Moore, R. J.; Angell, N. H.; Smith, R. D.; Springer, D. L.; Pounds, J. G. Toward a human blood serum proteome. *Mol. Cell. Proteomics* **2002**, *1*, 947–955.
- Anderson, N. L.; Anderson, N. G. The human plasma proteome: history, character, and diagnostic prospects. *Mol. Cell. Proteomics* **2002**, *1*, 845–867.
- Plaxco, K. W.; Gross, M. The importance of being unfolded. *Nature* **1997**, *386*, 657–659.
- Fink, A. L. Natively unfolded proteins. *Curr. Opin. Struct. Biol.* **2005**, *15*, 35–41.
- Kohn, J. E.; Plaxco, K. W. Engineering a signal transduction mechanism for protein-based biosensors. *Proc. Natl. Acad. Sci. U.S.A.* **2005**, *102*, 10841–10845.
- Oh, K. J.; Cash, K. C.; Plaxco, K. W. Beyond Molecular Beacons: Optical Sensors Based on the Binding-Induced Folding of Proteins and Polypeptides. *Chem.—Eur. J.* **2009**, *15*, 2244–2251.
- Stratton, M. M.; Mitrea, D. M.; Loh, S. N. A Ca²⁺-Sensing Molecular Switch Based on Alternate Frame Protein Folding. *ACS Chem. Biol.* **2008**, *3*, 723–732.
- Mok, W.; Li, Y. Recent progress in nucleic acid aptamer-based biosensors and bioassays. *Sensors* **2008**, *8*, 7050–7084.

- 28 Oh, K. J.; Cash, K. J.; Plaxco, K. W. Excimer-based peptide beacons: A convenient experimental approach for monitoring polypeptide–protein and polypeptide–oligonucleotide interactions. *J. Am. Chem. Soc.* **2006**, *128*, 14018–14019.
- 29 Oh, K. J.; Cash, K. J.; Hugenberg, V.; Plaxco, K. W. Peptide beacons: A new design for polypeptide-based optical biosensors. *Bioconjugate Chem.* **2007**, *18*, 607–609.
- 30 Oh, K. J.; Cash, K. J.; Lubin, A. A.; Plaxco, K. W. Chimeric peptide beacons: a direct polypeptide analog of DNA molecular beacons. *Chem. Commun.* **2007**, 4869–4871.
- 31 Tyagi, S.; Kramer, F. R. Molecular beacons: Probes that fluoresce upon hybridization. *Nat. Biotechnol.* **1996**, *14*, 303–308.
- 32 Thurley, S.; Roglin, L.; Seitz, O. Hairpin peptide beacon: Dual-labeled PNA-peptide-hybrids for protein detection. *J. Am. Chem. Soc.* **2007**, *129*, 12693–12695.
- 33 Fan, C. H.; Plaxco, K. W.; Heeger, A. J. Biosensors based on binding-modulated donor-acceptor distances. *Trends Biotechnol.* **2005**, *23*, 186–192.
- 34 Fan, C. H.; Plaxco, K. W.; Heeger, A. J. Electrochemical interrogation of conformational changes as a reagentless method for the sequence-specific detection of DNA. *Proc. Natl. Acad. Sci. U.S.A.* **2003**, *100*, 9134–9137.
- 35 Xiao, Y.; Lai, R. Y.; Plaxco, K. W. Preparation of electrode-immobilized, redox modified oligonucleotides for electrochemical DNA and aptamer-based sensing. *Nat. Protoc.* **2007**, *2*, 2875–2880.
- 36 Immoos, C. E.; Lee, S. J.; Grinstaff, M. W. Conformationally gated electrochemical gene detection. *ChemBioChem* **2004**, *5*, 1100–1103.
- 37 Mao, Y.; Luo, C.; Ouyang, Q. Studies of temperature-dependent electronic transduction on DNA hairpin loop sensor. *Nucleic Acids Res.* **2003**, *31*, e108.
- 38 Ricci, F.; Plaxco, K. W. E-DNA sensors for convenient, label-free electrochemical detection of hybridization. *Microchim. Acta* **2008**, *163*, 149–155.
- 39 Lee, T. M. Over-the-counter biosensors: past, present, and future. *Sensors* **2008**, *8*, 5535–5559.
- 40 Odenthal, K. J.; Gooding, J. J. An introduction to electrochemical DNA biosensors. *Analyst* **2007**, *132*, 603–610.
- 41 Mendes, P. M. Stimuli-responsive surfaces for bio-applications. *Chem. Soc. Rev.* **2008**, *37*, 2512–2529.
- 42 Privett, B. J.; Shin, J. H.; Schoenfish, M. H. Electrochemical sensors. *Anal. Chem.* **2008**, *80*, 4499–4517.
- 43 Lai, R. Y.; Lagally, E. T.; Lee, S. H.; Soh, H. T.; Plaxco, K. W.; Heeger, A. J. Rapid, sequence-specific detection of unpurified PCR amplicons via a reusable, electrochemical sensor. *Proc. Natl. Acad. Sci. U.S.A.* **2006**, *103*, 4017–4021.
- 44 Lubin, A. A.; Lai, R. Y.; Baker, B. R.; Heeger, A. J.; Plaxco, K. W. Sequence-specific, electronic detection of oligonucleotides in blood, soil, and foodstuffs with the reagentless, reusable E-DNA sensor. *Anal. Chem.* **2006**, *78*, 5671–5677.
- 45 Lubin, A. A.; Fan, C.; Schafer, M.; Clelland, C. T.; Bancroft, C.; Heeger, A. J.; Plaxco, K. W. Rapid Electronic Detection of DNA and Nonnatural DNA Analogs for Molecular Marking Applications. *Forensic Sci. Commun.* **2008**, *10* (1).
- 46 Lai, R. Y.; Seferos, D. S.; Heeger, A. J.; Bazan, G. C.; Plaxco, K. W. Comparison of the signaling and stability of electrochemical DNA sensors fabricated from 6- or 11-carbon self-assembled monolayers. *Langmuir* **2006**, *22*, 10796–10800.
- 47 Lai, R. Y.; Lee, S. H.; Soh, H. T.; Plaxco, K. W.; Heeger, A. J. Differential labeling of closely spaced biosensor electrodes via electrochemical lithography. *Langmuir* **2006**, *22*, 1932–1936.
- 48 Pavlovic, E.; Lai, R. Y.; Wu, T. T.; Ferguson, B. S.; Sun, R.; Plaxco, K. W.; Soh, H. T. Microfluidic device architecture for electrochemical patterning and detection of multiple DNA sequences. *Langmuir* **2008**, *24*, 1102–1107.
- 49 Lubin, A. A.; Vander Stoep Hunt, B.; White, R. J.; Plaxco, K. W. Effects of Probe Length, Probe Geometry, and Redox-Tag Placement on the Performance of the Electrochemical E-DNA Sensor. *Anal. Chem.* **2009**, *81*, 2150–2158.
- 50 Ricci, R.; Zari, N.; Caprio, F.; Recine, S.; Amine, A.; Moscone, D.; Paleschi, G.; Plaxco, K. W. Surface chemistry effects on the performance of an electrochemical DNA sensor. *Bioelectrochemistry* **2009**, *76*, 208–213.
- 51 Phares, N.; White, R. J.; Plaxco, K. W. Improving the stability and sensing of electrochemical biosensors by employing trithiol-anchoring groups in a six-carbon self-assembled monolayer. *Anal. Chem.* **2009**, *81*, 1095–1100.
- 52 Seferos, D. S.; Lai, R. Y.; Plaxco, K. W.; Bazan, G. C. α,ω -dithiol oligo(phenylene vinylene)s for the preparation of high-quality pi-conjugated self-assembled monolayers and nanoparticle-functionalized electrodes. *Adv. Funct. Mater.* **2006**, *16*, 2387–2392.
- 53 Ricci, F.; Lai, R. Y.; Heeger, A. J.; Plaxco, K. W.; Sumner, J. J. Effect of molecular crowding on the response of an electrochemical DNA sensor. *Langmuir* **2007**, *23*, 6827–6834.
- 54 Ricci, F.; Lai, R. Y.; Plaxco, K. W. Linear, redox modified DNA probes as electrochemical DNA sensors. *Chem. Commun.* **2007**, 3768–3770.
- 55 Ikeda, R.; Kobayashi, S.; Chiba, J.; Inoue, M. Detection of mismatched duplexes by synchronizing the pulse potential frequency with the dynamics of ferrocene/isoquinoline conjugate-connected DNA probes immobilized onto electrodes. *Chem.—Eur. J.* **2009**, *15*, 4822–4828.
- 56 Immoos, C. E.; Lee, S. J.; Grinstaff, M. W. DNA-PEG-DNA triblock macromolecules for reagentless DNA detection. *J. Am. Chem. Soc.* **2004**, *126*, 10814–10815.
- 57 Xiao, Y.; Lubin, A. A.; Baker, B. R.; Plaxco, K. W.; Heeger, A. J. Single-step electronic detection of femtomolar DNA by target-induced strand displacement in an electrode-bound duplex. *Proc. Natl. Acad. Sci. U.S.A.* **2006**, *103*, 16677–16680.
- 58 Xiao, Y.; Qu, X. G.; Plaxco, K. W.; Heeger, A. J. Label-free electrochemical detection of DNA in blood serum via target-induced resolution of an electrode-bound DNA pseudoknot. *J. Am. Chem. Soc.* **2007**, *129*, 11896–11897.
- 59 Cash, K. J.; Heeger, A. J.; Plaxco, K. W.; Xiao, Y. Optimization of a Reusable, DNA Pseudoknot-Based Electrochemical Sensor for Sequence-Specific DNA Detection in Blood Serum. *Anal. Chem.* **2009**, *81*, 656–661.
- 60 Ellington, A. D.; Szostak, J. W. In vitro selection of RNA molecules that bind specific ligands. *Nature* **1990**, *346*, 818–822.
- 61 Tuerk, C.; Gold, L. Systematic evolution of ligands by exponential enrichment - RNA ligands to bacteriophage-T4 DNA-polymerase. *Science* **1990**, *249*, 505–510.
- 62 Robertson, M. P.; Joyce, G. F. Selection in vitro of an RNA enzyme that specifically cleaves single-stranded-DNA. *Nature* **1990**, *344*, 467–470.
- 63 Xiao, Y.; Plaxco, K. W. Electrochemical Approaches to Aptamer-Based Sensing. In *Functional Nucleic Acids for Analytical Applications*; Lu, Y., Li, Y., Eds.; Springer Science + Business Media: New York, 2009; Chapter 7.
- 64 Hamaguchi, N.; Ellington, A.; Stanton, M. Aptamer beacons for the direct detection of proteins. *Anal. Biochem.* **2001**, *294*, 126–131.
- 65 Li, J. W. J.; Fang, X. H.; Tan, W. H. Molecular aptamer beacons for real-time protein recognition. *Biochem. Biophys. Res. Commun.* **2002**, *292*, 31–40.
- 66 Potyrailo, R. A.; Conrad, R. C.; Ellington, A. D.; Hieftje, G. M. Adapting selected nucleic acid ligands (aptamers) to biosensors. *Anal. Chem.* **1998**, *70*, 3419–3425.
- 67 Baldrich, E.; Restrepo, A.; O'Sullivan, C. K. Aptasensor development: elucidation of critical parameters for optimal aptamer performance. *Anal. Chem.* **2004**, *76*, 7053–7063.
- 68 Bock, L. C.; Griffin, L. C.; Latham, J. A.; Vermaas, E. H.; Toole, J. J. Selection of single-stranded DNA molecules that bind and inhibit human thrombin. *Nature* **1992**, *355*, 564–566.
- 69 Dittmer, W. U.; Reuter, A.; Simmel, F. C. A DNA-based machine that can cyclically bind and release thrombin. *Angew. Chem., Int. Ed.* **2004**, *43*, 3550–3553.
- 70 Macaya, R. F.; Schultze, P.; Smith, F. W.; Roe, J. A.; Feigon, J. Thrombin-binding DNA aptamer forms a unimolecular quadruplex structure in solution. *Proc. Natl. Acad. Sci. U.S.A.* **1993**, *90*, 3745–3749.
- 71 Xiao, Y.; Lubin, A. A.; Heeger, A. J.; Plaxco, K. W. Label-free electronic detection of thrombin in blood serum by using an aptamer-based sensor. *Angew. Chem., Int. Ed.* **2005**, *44*, 5456–5459.
- 72 Xiao, Y.; Piorek, B. D.; Plaxco, K. W.; Heeger, A. J. A reagentless signal-on architecture for electronic, aptamer-based sensors via target-induced strand displacement. *J. Am. Chem. Soc.* **2005**, *127*, 17990–17991.
- 73 Radi, A.; Sanchez, J. L. A.; Baldrich, E.; O'Sullivan, C. K. Reagentless, reusable, ultrasensitive electrochemical molecular beacon aptasensor. *J. Am. Chem. Soc.* **2006**, *128*, 117–124.
- 74 Sanchez, J. L. A.; Baldrich, E.; Radi, A. E.; Dondapati, S.; Sanchez, P. L.; Katakis, I.; O'Sullivan, C. K. Electronic 'Off-On' molecular switch for rapid detection of thrombin. *Electroanalysis* **2006**, *18*, 1957–1962.
- 75 Baker, B. R.; Lai, R. Y.; Wood, M. S.; Doctor, E. H.; Heeger, A. J.; Plaxco, K. W. An electronic, aptamer-based small-molecule sensor for the rapid, label-free detection of cocaine in adulterated samples and biological fluids. *J. Am. Chem. Soc.* **2006**, *128*, 3138–3139.
- 76 White, R. J.; Phares, N.; Lubin, A. A.; Xiao, Y.; Plaxco, K. W. Optimization of electrochemical aptamer-based sensors via optimization of probe packing density and surface chemistry. *Langmuir* **2008**, *24*, 10513–10518.
- 77 Swensen, J. S.; Xiao, Y.; Ferguson, B. S.; Lubin, A. A.; Lai, R. Y.; Heeger, A. J.; Plaxco, K. W.; Soh, H. T. Continuous, real-time monitoring of cocaine in undiluted blood serum via a microfluidic, electrochemical aptamer-based sensor. *J. Am. Chem. Soc.* **2009**, *131*, 4262–4266.
- 78 Radi, A. E.; O'Sullivan, C. K. Aptamer conformational switch as sensitive electrochemical biosensor for potassium ion recognition. *Chem. Commun.* **2006**, 3432–3434.
- 79 Lai, R. Y.; Plaxco, K. W.; Heeger, A. J. Aptamer-based electrochemical detection of picomolar platelet-derived growth factor directly in blood serum. *Anal. Chem.* **2007**, *79*, 229–233.

- 80 Peng, Y.; Wang, X.; Xiao, Y.; Feng, L.; Zhao, C.; Ren, J.; Qu, X. i-Motif quadruplex DNA-based biosensor for distinguishing single- and multiwalled carbon nanotubes. *J. Am. Chem. Soc.* **2009**, *131*, 13813–13818.
- 81 Han, D.; Kim, Y.; Oh, J.; Kim, T. H.; Mahajan, R. K.; Kim, J. S.; Kim, H. A regenerative electrochemical sensor based on oligonucleotide for the selective determination of mercury (II). *Analyst* **2009**, *134*, 1857–1862.
- 82 Ferapontova, E. E.; Olsen, E. M.; Gothelf, K. V. An RNA aptamer-based electrochemical biosensor for detection of theophylline in serum. *J. Am. Chem. Soc.* **2008**, *130*, 4256–4258.
- 83 Ferapontova, E. E.; Gothelf, K. V. Effect of serum on an RNA aptamer-based electrochemical sensor of theophylline. *Langmuir* **2009**, *25*, 4279–4283.
- 84 Ferapontova, E. E.; Gothelf, K. V. Optimization of the electrochemical RNA-aptamer based biosensor for theophylline by using a methylene blue redox label. *Electroanalysis* **2009**, *21*, 1261–1266.
- 85 Lu, Y.; Li, X.; Zhang, L.; Yu, P.; Su, L.; Mao, L. Aptamer-based electrochemical sensors with aptamer-complementary DNA oligonucleotides as probe. *Anal. Chem.* **2008**, *80*, 1883–1890.
- 86 Zuo, X.; Song, S.; Zhang, J.; Pan, D.; Wang, L.; Fan, C. A target-responsive electrochemical aptamer switch (TREAS) for reagentless detection of nanomolar ATP. *J. Am. Chem. Soc.* **2007**, *129*, 1042–1043.
- 87 Xiao, Y.; Rowe, A. A.; Plaxco, K. W. Electrochemical detection of parts-per-billion lead via an electrode-bound DNAzyme assembly. *J. Am. Chem. Soc.* **2007**, *129*, 262–263.
- 88 Ricci, F.; Bonham, A. J.; Mason, A. C.; Reich, N. O.; Plaxco, K. W. Reagentless, electrochemical approach for the specific detection of double- and single-stranded DNA binding proteins. *Anal. Chem.* **2009**, *81*, 1608–1614.
- 89 Cash, K. J.; Ricci, F.; Plaxco, K. W. An electrochemical sensor for the detection of protein-small molecule interactions directly in serum and other complex matrices. *J. Am. Chem. Soc.* **2009**, *131*, 6955–6957.
- 90 Cash, K.; Ricci, F.; Plaxco, K. A general electrochemical method for label-free screening of protein-small molecule interactions. *Chem. Commun.* **2009**, (41), 6222–6224. DOI = 10.1039/b911558g.